

REV: 7/10

## PRODUCT INFORMATION SUMMARY

### Human IL-6 "SUPER X" ELISA Kit

Product Number **RH600020CKX**  
5 Plate Kit ( 5 x 96 tests )  
**RH600020CKX2** ( 2 x 96 tests )  
**RH600020CKX1** ( 1 x 96 tests )

#### Items Provided:

**PRE-COATED ELISA Plates** ( Single, 2 or 5 plates)  
Biotin-Labeled tracer 25.0 ug **or** 6.0 ug ( **see vial**)  
Antigen Standard 1.0 ug **or** 5.0 ug ( **see vial**)  
Streptavidin-HRP 1.0 mL ( 0.5 mL - 2 plate format)  
TMB Substrate 25 mL x 2 ( 15 mL - 2 plate)  
Wash Buffer Concentrate 125 mL ( 20X) ( 50 mL - 2 plate)

#### DESCRIPTION:

Human Interleukin 6 (IL-6) **SUPER X ELISA Kit** provides one, two or five **PRE-COATED** microplates ( coated with antigen affinity purified capture antibody stabilized with our proprietary ELISA coating/blocking reagent. A biotin labeled tracer antibody, antigen standard, HRP developing reagents and wash buffer are included.

Final working concentrations and assay range may need to be optimized by customer. Reactivity with various sample types, including serum/plasma samples, is evaluated by customer.

Note: Reconstitute components only when ready to run assay.

#### TRACER ANTIBODY:

Provided as 25 ug lyophilized or 0.5 mL liquid ( see vial) of Biotin labeled, antigen-affinity purified antibody, additive-free. For 25.0 ug lyophilized : Reconstitute in 200 uL sterile water **containing 0.1% BSA.** ( FREEZE aliquots for long-term storage)

For liquid vial ( 0.5 mL) - store refrigerated only.

**STANDARD:** Provided as 1.0 ug **or** 5.0 ug ( see vial) of recombinant Human IL-6. Quick-spin and **reconstitute in distilled water** (pH 8.0) - concentration approx. 100.0 ug/mL. Further dilutions can be made in 0.05% Tween-20, 0.1% BSA in PBS.

**DEVELOPING REAGENTS:**

- Streptavidin-HRP ( S100180CX) 1.0 mL - **store @ -20 Deg. C.**
- TMB Substrate Solutions - Part A and Part B ( 25.0 mL each) cat # ES200CX
- Wash Buffer Concentrate (20X concentrate) Mix 1 volume of wash buffer with 19 volumes of distilled water. Stable for one month @ 4 Deg. C. once mixed to working volume. Concentrate may be stored at room temp.

**HANDLING/ STORAGE:** Reconstitute reagents when ready to build ELISA assay. Reconstituted Tracer Antibody be stored for approximately one month at 4 Degrees C. Or store **frozen** at -20 Degrees C. for up to 6 months.

For tracer antibody ( 0.5 mL) provided liquid- store refrigerated only- contains preservative.

Standard ( rec. IL-6) can be stored in liquid state ( @ 4 Deg. C.) For up to one week, or store **frozen, with addition of 0.1% BSA**, at -20 Deg. C. for up to 2 months. AVOID repeat freeze-thaw. Pre-coated ELISA plates should be stored refrigerated ( @ 4 Deg.C.) in sealed plastic bag with desiccant pack, and are stable for one year from date of receipt. If needed, additional components ( antigen standard or developing reagents) can be purchased as separate items to run additional stored plates - inquire.

**Store** Streptavidin-HRP ( S100180XH) **frozen** @ -20 Deg. C.

Store TMB solutions and wash buffer concentrate at 4 Deg. C.  
**CAUTION: Substrate Solution B contains 20% acetone. FLAMMABLE.**  
**Keep away from sources of heat or flame.**

**MATERIALS RECOMMENDED:**

Tween -20.  
BSA ( ELISA grade only, protease-free)  
Dubelco's PBS (10X)  
PBS: Dilute to 1XPBS in sterile water  
**Diluent:** use : 0.05% Tween-20, 0.1% BSA in PBS  
2N Sulfuric acid ( stop solution).

**PLATE PREPARATION:**

Plates are **pre-coated** with capture antibody and blocked/stabilized with ANTIGENIX proprietary ELISA coating stabilizer ( EA150) and are **ready to use!**

Store plates refrigerated ( @ 4 Deg. C.) in sealed plastic bags with desiccant pack. Plates can be stored for one year from data of receipt.

**PROTOCOL:**

**STANDARD/SAMPLE:** Dilute a portion of the standard ( store unused standard in aliquots, high concentration, frozen -20 Deg. C. with addition of 0.1% BSA) from **2.0 ng/mL** (**Adjust**, depending on desired range and sensitivity of first standard curve) to zero in diluent (serial dilution). Immediately add 100 uL of standard or sample to each well in duplicate. **Incubate** at room temp. for **1 hour**.

**DETECTION:** Aspirate and wash plate 4 times. **Dilute** detection (Biotin Tracer) antibody in diluent to **concentration** of **0.10 ug/mL**. Add 100 uL per well. Incubate at room temperature for **1 hour**.

**STREPTAVIDIN-HRP:** Aspirate and wash plate 4 times. **Dilute** Streptavidin-HRP conjugate approx. **1:3,000** in diluent. (May need to optimize) Add 100 uL per well, incubate 30 minutes at room temperature.

**SUBSTRATE:** **Prepare substrate solution** no more than 15 minutes before last incubation of assay: Mix one part TMB Solution A with one part TMB Solution B in a clean container. If, upon mixing, TMB solution turns blue - TMB solution is **contaminated-DO NOT USE**. Use mixed substrate solution **WITHIN 2 Hours, and AVOID DIRECT LIGHT**.

Aspirate and wash plate 4 times. **Note:** Wash steps are critical! Add 100 uL substrate solution to each well. **Incubate** at room temp. for color development. Monitor color development with plate reader at 650 nm wavelength. ( for blue color). **Add** 100 uL of **Stop solution** ( 2N Sulfuric Acid) **within 10 minutes** to stop color development - gently tap plate to mix. **Read** plate at **450 nm** within 30 minutes of addition of stop solution.

**NOTE:** Plate may develop with TMB in as little as **5-10 minutes!**

**NOTE:** reliable standard curves are obtained when O.D. readings do not exceed 0.25 units for the zero standard concentration, or 2.0 units for the highest standard concentration. Monitor the plate every 5 minutes for approximately 30 minutes.

**WARRANTY:**

Products sold hereunder are warranted only to conform to the quantity and contents stated on the label at the time of delivery to the customer. There are no warranties, expressed or implied, which extend beyond the description on the label of the product.

**RESEARCH USE ONLY -NOT For DIAGNOSTIC USE**

ANTIGENIX AMERICA Inc.

[WWW.ANTIGENIX.COM](http://WWW.ANTIGENIX.COM)