

## PRODUCT INFORMATION SUMMARY

### Human Endostatin

Product Number **RHF503CKX**

### "SUPER X" ELISA Kit

5 Plate Kit ( 5 x 96 tests )  
**RHF503CKX2** ( 2 Plate Kit )  
**RHF503CKX1** ( Single Plate Kit )

#### Items Provided:

#### **PRE-COATED ELISA Plates** (Single, 2 or 5 plates)

Biotin-Labeled tracer	25.0 ug
Antigen Standard	20.0 ug
Streptavidin-HRP	1.0 mL ( 0.5 mL - 2 plate format)
TMB Substrate	25 mL x 2 ( 15 mL - 2 plate)
Wash Buffer Concentrate	125 mL ( 20X) ( 50 mL - 2 plate)

#### **DESCRIPTION:**

Human Endostatin **SUPER X ELISA Kit** provides one, two or five **PRE-COATED** microplates ( coated with antigen affinity purified capture antibody stabilized with our proprietary ELISA coating/blocking reagent. A biotin labeled tracer antibody, antigen standard, HRP developing reagents and wash buffer are included.

Final working concentrations and assay range may need to be optimized by customer. Reactivity with various sample types, including serum/plasma samples, is evaluated by customer.

Note: Reconstitute components only when ready to run assay.

#### **TRACER ANTIBODY:**

Provided as 25 ug of Biotin labeled, antigen-affinity purified antibody, additive-free. Reconstitute in 500 uL sterile water **containing 0.1% BSA.** ( FREEZE aliquots for long-term storage)

**STANDARD:** Provided as 20.0 ug of recombinant Endostatin. Quick-spin and **reconstitute in distilled water** (pH 8.0) - concentration approx. 0.1 mg/mL. Further dilutions can be made in 0.05% Tween-20, 0.1% BSA in PBS.

#### **DEVELOPING REAGENTS:**

- Streptavidin-HRP ( S100180CX) 1.0 mL - **store @ -20 Deg. C.**
- TMB Substrate Solutions - Part A and Part B ( 25.0 mL each) cat # ES200CX
- Wash Buffer Concentrate (20X concentrate) Mix 1 volume of wash buffer with 19 volumes of distilled water. Stable for one month @ 4 Deg. C. once mixed to working volume.

**HANDLING/ STORAGE:** Reconstitute reagents when ready to build ELISA assay. Tracer Antibody be stored for approximately one month at 4 Degrees C. Or store **frozen** at -20 Degrees C. for up to 6 months. Standard ( rec. Endostatin) can be stored in liquid state ( @ 4 Deg. C.) For up to one week, or store **frozen, with addition of 0.1% BSA**, at -20 Deg. C. for up to 2 months. AVOID repeat freeze-thaw. Pre-coated ELISA plates should be stored refrigerated ( @ 4 Deg.C.) in sealed plastic bag with desiccant pack, and are stable for one year from date of receipt. If needed, additional components ( antigen standard or developing reagents) can be purchased as separate items to run additional stored plates - inquire.

**Store** Streptavidin-HRP ( S100180CX) **frozen** @ -20 Deg. C.

Store TMB solutions and wash buffer concentrate at 4 Deg. C.

**CAUTION: Substrate Solution B contains 20% acetone. FLAMMABLE.**  
**Keep away from sources of heat or flame.**

**MATERIALS RECOMMENDED:**

Tween -20.

BSA ( ELISA grade only, protease-free)

Dubelco's PBS (10X)

PBS: Dilute to 1XPBS in sterile water

**Diluent:** use : 0.05% Tween-20, 0.1% BSA in PBS

2N Sulfuric acid - Stop Solution.

**PLATE PREPARATION:**

Plates are **pre-coated** with capture antibody and blocked/stabilized with ANTIGENIX proprietary ELISA coating stabilizer ( EA150) and are **ready to use!**

Store plates refrigerated ( @ 4 Deg. C.) in sealed plastic bags with desiccant pack. Plates can be stored for one year from data of receipt.

**PROTOCOL:**

**STANDARD/SAMPLE:** Dilute **a portion of the** standard ( store unused standard in aliquots, high concentration, frozen -20 Deg. C. with addition of 0.1% BSA) from **4.0 ng/mL** to zero in diluent (serial dilution). Immediately add 100 uL of standard or sample to each well in duplicate. **Incubate** at room temp. for **1 hour**.

**DETECTION:** Aspirate and wash plate 4 times. Dilute portion of detection (Biotin Tracer) antibody in diluent to concentration of **0.10 ug/mL**. Add 100 uL per well. Incubate at room

temperature for **30 minutes**.

**STREPTAVIDIN-HRP:** Aspirate and wash plate 4 times. **Dilute** Streptavidin-HRP conjugate approx. **1:2,500** in diluent. (May need to optimize) Add 100 uL per well, incubate 30 minutes at room temperature.

**SUBSTRATE: Prepare substrate solution** no more than 15 minutes before last incubation of assay: Mix one part TMB Solution A with one part TMB Solution B in a clean container. If, upon mixing, TMB solution turns blue - TMB solution is **contaminated-DO NOT USE**. Use mixed substrate solution **WITHIN 2 Hours, and AVOID DIRECT LIGHT**.

Aspirate and wash plate 4 times. **Note:** Wash steps are critical! Add 100 uL substrate solution to each well. Incubate at room temp. for color development. **Add 100 uL of Stop solution** ( 2N Sulfuric Acid) within **10-15 minutes** to stop color development - gently tap plate to mix. **Read** plate at **450 nm** within 30 minutes of addition of stop solution.

**NOTE:** reliable standard curves are obtained when O.D. readings do not exceed 0.35 units for the zero standard concentration, or 1.5 units for the highest standard concentration.

**RESEARCH USE ONLY -NOT For DIAGNOSTIC USE**

**NOTE:** Addition components may be ordered separately, as needed, if running additional stored plates in future:

**Super X ELISA Kit**

**Accessory Reagents Available:**

Additional antigen standard (\$100.) Or other components may be purchased as separate item- Inquire.

ANTIGENIX AMERICA Inc.

[WWW.ANTIGENIX.com](http://WWW.ANTIGENIX.com)