#### ANTIGENIX AMERICA Inc.

# Super-X Plex<sup>™</sup> Flow Cytometry Assay Multi-Plex Panel (6 Plex)

Product Information Summary

## Human Th1/Th2 - Panel #2 Cat No.HMX786T 32 tests

DESCRIPTION: Multi-Plex ( 6 Plex) Flow Cytometry Assay Panel
- components to measure human: IL-2, IL-4, IL-5, IL-10,
IFN Gamma, and TNF Alpha simultaneously- in human Serum,
Plasma, Cell Culture Supernatant or Cell Lysate samples.

NOTE: ORDER a Sample Diluent Kit ( to match sample type):

Human Serum/Plasma (**DKH100**) or Cell Culture Supernatant (**DKC100**) or Cell Lysate (**DKL100**)

Cell Lysate Prep Buffer (CLP200) or Sample Dilution Buffer (to dilute high concentration samples) - (SDL200) also available.

#### (STARTER Kit Included)

#### CONTENTS:

Antibody Conjugated Beads 1.5 mL (1X) (Pre-Mixed , 6-Plex)
Detection Antibody - Biotin 0.45 mL (2X) (Pre-Mixed , 6-Plex)
Detection Antibody Diluent 0.45 mL (2X)

Standard: One vial of lyophilized standard containing

**Pre-mixed multiplexed** analytes ( see vial insert) -single standard vial can be used for one or more assays indicated

#### STARTER Kit: ( Human Panel)

Reading Buffer 5.0 mL (10X)
Wash Buffer 15.0 mL (10X)
Filter plate with lid (96 wells)
Plate Sealers (6)
Streptavidin-PE 3.0 mL (1X)
PCR 8 tube strip 2 strips

STORAGE: Store Refrigerated ( 4-8°C. - in dark)

CAUTION: Contains Sodium Azide - ( 0.05%) - Dispose contents carefully, flush with water to avoid reaction with lead drain lines.

**PROTOCOL:** Refer to Kit protocol provided with assay ordered. PROTOCOL can also be DOWNLOADED @ www.antigenix.com

#### Assay Specifications:

Bead	Specs:	8	Sensitiviti	Les	(pg/mL	)	Ra	inge :	(pg/mL)
	S4P3 S4P7 S4P9 S5P4 S5P5 S5P10	Human Human Human	IFN Gamma IL-4 IL-5	<	3.0	10.0 10.0 3.0 5.0	- - -	1,000 2,000 2,000 1,000 5,000 5,000	

Sample Size: 15.0 uL per test

X-Reactivity within panel : negligable

Intra Assay CV: <10% Inter Assay CV <20%

Recovery 70-130% standard dose

### RESEARCH USE ONLY NOT INTENDED FOR DIAGNOSTIC OR THERAPEUTIC USE.

WWW.ANTIGENIX.com 1-800-558-1008