

PRODUCT INFORMATION SUMMARY

Human sol.CD22 ELISA Construction Kit

Product Number RHF225CK
Approx. 960 tests

Product Number RHF225CKC
With Developing Reagents

Capture Antibody	50.0 ug	ELISA Coating Stabilizer	50mL
Biotin-Labeled tracer	25.0 ug	Streptavidin-HRP	1.0mL
Antigen Standard	5.0 ug	TMB Substrate	(50 ML x 2)

DESCRIPTION:

Human Soluble CD22 is a 149 aa. protein with molecular weight of 75.0 kDa. This ELISA CONSTRUCTION Kit provides antigen affinity purified polyclonal capture and tracer antibodies, and antigen standard sufficient for **approximately** five to ten microplates.

Working concentrations must be optimized by customer.

Note: Reconstitute components only when ready to run assay.

CAPTURE ANTIBODY:

Provided as lyophilized, 50 ug, additive-free. Reconstitute in 0.50 mL sterile water (0.1 mg/mL). (FREEZE aliquots for long-term storage)

TRACER ANTIBODY:

Provided as 25 ug of Biotin labeled, antigen-affinity purified antibody, additive-free. Reconstitute in 500 uL sterile water **containing 0.1% BSA**. (FREEZE aliquots for long-term storage)

STANDARD: Provided as 5.0 ug of recombinant Human sCD22. Quick-spin and reconstitute in 100 uL of 10 mM NaP (pH 7.2). Further dilutions can be made in 0.05% Tween-20, 0.1% BSA in PBS.

DEVELOPING REAGENTS: Supplied with catalog # ending in "CKC".

- ELISA Coating/ Blocking Reagent (EA150C) 50.0 mL (5X Solution)
- Streptavidin-HRP (S100180C) 1.0 mL - store @ -20 Deg. C.
- TMB Substrate Solutions - Part A and Part B (50.0 mL each)
cat # ES200C

HANDLING/ STORAGE: Reconstitute reagents when ready to build ELISA assay. Antibodies (Capture and Tracer) can be stored for approximately one month at 4 Degrees C. Or store **frozen** at -20 Degrees C. for up to 6 months. Standard (rec. sCD22) can be stored in liquid state (@ 4 Deg. C.) For up to one week, or store **frozen** at -20 Deg. C. for up to 2 months. AVOID repeat freeze-thaw.

MATERIALS RECOMMENDED:

ELISA Microplates: Nunc Maxisorp, Prod. # 4420404
Tween -20.
BSA
Streptavidin-HRP: ANTIGENIX Cat no. S100180 or similar
TMB Substrate
Dubelco's PBS (10X)
ANTIGENIX ELISA Coating Stabilizer (cat no: EA150)

RECOMMENDED SOLUTIONS:

PBS: Dilute to 1XPBS in sterile water
WASH BUFFER: 0.05% Tween-20 in PBS.
BLOCK BUFFER: **use ANTIGENIX AMERICA coating stabilizer (EA150) or 1% BSA in PBS**
Substrate Solution: TMB Substrate Solution
Diluent: 0.05% Tween-20, 0.1% BSA in PBS
2N Sulfuric Acid (Stop Solution)

PLATE PREPARATION:

1. Dilute **100 uL of** capture antibody with 0.05M Carbonate buffer (or PBS) to concentration 1.0 ug/mL. To optimize coating concentration on first test plate, use 3 test concentrations of capture antibody (for example: 1.0 ug/mL, 0.5 ug/mL and 0.25 ug/mL).
Immediately add 100 uL to each ELISA well. (With first test plate, coat one-third of plate with each concentration, and note which wells have different concentrations). Seal the plate and incubate overnight at room temperature.
2. Aspirate wells to remove all liquid and wash **4 times** using 300 uL of wash buffer per well. After last wash, add 200 uL ANTIGENIX AMERICA **ELISA coating stabilizer (cat # EA150)** and incubate for 60 minutes at room temperature. (With coating stabilizer, DO NOT let plate dry prior to use of coating stabilizer. This will **stabilize and Block in one step!** Refer to data sheet EA150 for complete description of use.
3. With ANTIGENIX coating stabilizer (**recommended**) aspirate plate but **DO NOT WASH**. Dry plate in humidity controlled chamber or similar. (see data sheet EA150). With standard block reagent, aspirate plate and wash 3X with 300 uL wash buffer

PROTOCOL:

STANDARD/SAMPLE: Dilute **a portion of the** standard (store unused standard in aliquots, high concentration, frozen -20 Deg. C.) from 100.0 ng/mL (depending on desired range) to zero in diluent (serial dilution). Immediately add 100 uL of standard or sample to each well in duplicate. Incubate at room temp. for approx. 2 hours.

DETECTION: Aspirate and wash plate 4 times. Dilute portion of detection (Tracer) antibody in diluent to concentration of 0.20 ug/mL. Add 100 uL per well. Incubate at room temperature for 2 hours. Note: detection antibody can be used in approximate range of 0.10 - 0.50 ug/mL, you may need to optimize for subsequent plates.

STREPTAVIDIN-HRP: Aspirate and wash plate 4 times. Dilute Streptavidin-HRP conjugate approx. 1:.2,000 in diluent (follow recommended dilution of manufacturer). (May need to optimize) Add 100 uL per well, incubate 30 minutes at room temperature.

SUBSTRATE: Aspirate and wash plate 4 times. Add 100 uL substrate solution to each well. Incubate at room temp. for color development. To monitor color development- set plate reader at 650 nm wavelength (blue color - prior to addition of stop solution). The color reaction may be stopped after 10 - 30 minutes by adding 100 uL of 2N Sulfuric acid to each well. Then, read plate at 450 nm with correction set to 650 nm) after reaction has been stopped.

NOTE: reliable standard curves are obtained when O.D. readings do not exceed 0.2 units for the zero standard concentration, or 2.0 units for the highest standard concentration. Monitor the plate every 5 minutes for approximately 30 minutes.

WARRANTY:

Products sold hereunder are warranted only to conform to the quantity and contents stated on the label at the time of delivery to the customer. There are no warranties, expressed or implied, which extend beyond the description on the label of the product.

RESEARCH USE ONLY -NOT For DIAGNOSTIC USE

NOTE: Kit can be ordered with the following developing reagents (suitable for approx. 1,000 wells - ten microplates) or order separately as below:

ELISA CONSTRUCTION KITS

Accessory Reagents Available:

Streptavidin-HRP; S100180, 1.0 mL, Suitable for ~5,000 ELISA wells, \$190.00

TMB Substrate; ES200, 100 mL x 2; Suitable for ~2,000 ELISA wells,\$95.00

ELISA Coating Stabilizer; EA150, 100 mL (5X); Suitable for ~ 2,500 wells
\$190.00

Get all three reagents above as "ELISA Construction Pack" ; EA700, \$410.00

ANTIGENIX AMERICA Inc.

WWW.ANTIGENIX.com

