

## PRODUCT INFORMATION SUMMARY

### Human MIG

Product Number **RHF450CKX**

### "SUPER X" ELISA Kit

5 Plate Kit ( 5 x 96 tests)

Product Number **RHF450CKX2** (2 x 96 tests)

Product Number **RHF450CKX1** ( 1 x 96 tests )

#### Items Provided:

**PRE-COATED ELISA Plates** (5 plates, 2 plates or single plate)  
Biotin-Labeled tracer 0.5 mL or 25.0 ug (**see vial**)  
Antigen Standard 5.0 ug  
Streptavidin-HRP 0.5 mL  
TMB Substrate 25 mL x 2 ( 15.0 mL - 2 plate)  
Wash Buffer Concentrate 125 mL ( 20X); or 50 mL ( 2 plate kit)

#### DESCRIPTION:

Human Monokine Induced by Interferon Gamma (MIG) **SUPER X ELISA Kit** provides one, two or five **PRE-COATED** microplates ( coated with antigen affinity purified capture antibody stabilized with our proprietary ELISA coating/blocking reagent. A biotin labeled tracer antibody, antigen standard, HRP developing reagents and wash buffer are included.

Final working concentrations and assay range may need to be optimized by customer. Reactivity with various sample types, including serum/plasma samples, is evaluated by customer.

Note: Reconstitute components only when ready to run assay.

#### TRACER ANTIBODY:

Provided as 25 ug ( lyophilized )or as 0.5 mL liquid @ 33.0 ug/mL ( **see vial**) of Biotin labeled, antigen-affinity purified antibody, additive-free. Reconstitute the 25 ug ( lyophilized vial) in 500 uL sterile water **containing 0.1% BSA**. ( FREEZE aliquots for long-term storage). -

**\*\*For liquid vial store refrigerated only ( contains preservative)\*\*.**

**STANDARD:** Provided as 5.0 ug of recombinant Human MIG. Quick-spin and **reconstitute in distilled water** (pH 8.0) - concentration approx. 0.1 mg/mL. Further dilutions can be made in 0.05% Tween-20, 0.1% BSA in PBS.

#### DEVELOPING REAGENTS:

- Streptavidin-HRP ( S100180CX) 1.0 mL - **store @ -20 Deg. C.**
- TMB Substrate Solutions - Part A and Part B ( 25.0 mL or 15.0 mL each) cat # ES200CX
- Wash Buffer Concentrate (20X concentrate) Mix 1 volume of wash buffer with 19 volumes of distilled water. Stable for one month

@ 4 Deg. C. once mixed to working volume.

**HANDLING/ STORAGE:** Reconstitute reagents when ready to build ELISA assay. Biotin Tracer Antibody provided as 25.0 ug lyophilized can be stored for approximately one month at 4 Degrees C. Or store **frozen** at -20 Degrees C. for up to 6 months.

Biotin Tracer antibody provided as 0.5mL liquid ( **see vial**)-  
STORE REFRIGERATED only - contains preservative.

Standard ( rec. MIG) can be stored in liquid state ( @ 4 Deg. C.)  
For up to one week, or store **frozen, with addition of 0.1% BSA**, at  
-20 Deg. C. for up to 2 months. AVOID repeat freeze-thaw. Pre-  
coated ELISA plates should be stored in sealed plastic bag with  
desiccant pack, and are stable for one year from date of receipt.  
If needed, additional components ( antigen standard or developing  
reagents) can be purchased as separate items to run additional  
stored plates - inquire.

**Store** Streptavidin-HRP ( S100180XH) **frozen** @ -20 Deg. C.

Store TMB solutions and wash buffer concentrate at 4 Deg. C.  
**CAUTION: Substrate Solution B contains 20% acetone. FLAMMABLE.**  
**Keep away from sources of heat or flame.**

#### **MATERIALS RECOMMENDED:**

Tween -20.  
BSA ( ELISA grade only, protease-free)  
Dubelco's PBS (10X)  
PBS: Dilute to 1XPBS in sterile water  
**Diluent:** use : 0.05% Tween-20, 0.1% BSA in PBS

#### **PLATE PREPARATION:**

Plates are **pre-coated** with capture antibody and  
blocked/stabilized with ANTIGENIX proprietary ELISA coating  
stabilizer ( EA150) and are **ready to use!**

Store plates refrigerated ( @ 4 Deg. C.) in sealed plastic bags  
with desiccant pack. Plates can be stored for one year from  
data of receipt.

#### **PROTOCOL:**

**STANDARD/SAMPLE:** Dilute **a portion of the** standard ( store unused  
standard in aliquots, high concentration, frozen -20 Deg. C.  
with addition of 0.1% BSA) from **10.0 ng/mL** to zero in diluent  
(serial dilution). Immediately add 100 uL of standard or sample  
to each well in duplicate. **Incubate** at room temp. for **1 hour**.

**DETECTION:** Aspirate and wash plate 4 times. Dilute detection  
(Biotin Tracer) antibody in diluent to **concentration** of

**0.25 ug/mL.** Add 100 uL per well. **Incubate** at room temperature for **1 hour**.

**STREPTAVIDIN-HRP:** Aspirate and wash plate 4 times. **Dilute** Streptavidin-HRP conjugate approx. **1:2,500** in diluent. (May need to optimize) Add 100 uL per well, **incubate 30 minutes** at room temperature.

**SUBSTRATE: Prepare substrate solution** no more than 15 minutes before last incubation of assay: Mix one part TMB Solution A with one part TMB Solution B in a clean container. If, upon mixing, TMB solution turns blue - TMB solution is **contaminated-DO NOT USE**. Use mixed substrate solution **WITHIN 2 Hours, and AVOID DIRECT LIGHT**.

Aspirate and wash plate 4 times. **Note:** Wash steps are critical! **Add** 100 uL **substrate solution** to each well. Incubate at room temp. for color development. Monitor color development with plate reader at 650 nm wavelength. ( for blue color). **Add** 100 uL of Stop solution ( 2N Sulfuric Acid) **within 15-20 minutes** to stop color development - gently tap plate to mix. Read plate at 450 nm within 30 minutes of addition of stop solution.

NOTE: reliable standard curves are obtained when O.D. readings do not exceed 0.25 units for the zero standard concentration, or 2.0 units for the highest standard concentration.

**WARRANTY:**

Products sold hereunder are warranted only to conform to the quantity and contents stated on the label at the time of delivery to the customer. There are no warranties, expressed or implied, which extend beyond the description on the label of the product.

**RESEARCH USE ONLY -NOT For DIAGNOSTIC USE**

ANTIGENIX AMERICA Inc.

[WWW.ANTIGENIX.com](http://WWW.ANTIGENIX.com)