PRODUCT INFORMATION SUMMARY

Human TGF Alpha ELISA Construction Kit

Product Number RHF530CK Approx. 960 tests		Product Number RHF530CKC With Developing Reagents:	:
Capture Antibody	50.0 ug	ELISA Coating Stabilizer	50 mL
Biotin-Labeled tracer	25.0 ug	Streptavidin-HRP	1.0mL
Antigen Standard	20.0 ug	TMB Substrate (50 mL x	2)

DESCRIPTION:

This ELISA CONSTRUCTION Kit provides antigen affinity purified polyclonal capture and tracer antibodies, and antigen standard sufficient for **approximately** five to ten microplates. Working concentrations must be optimized by customer.

Note: Reconstitute components only when ready to run assay.

CAPTURE ANTIBODY:

Provided as lyophilized, 50 ug, additive-free. Reconstitute in 0.50 mL sterile water (0.1 mg/mL). (<code>FREEZE</code> aloquots for long-term storage)

TRACER ANTIBODY:

Provided as 25 ug of Biotin labeled, antigen-affinity purified antibody, additive-free. Reconstitute in 500 uL sterile water containing 0.1% BSA. (FREEZE aloquots for long-term storage)

STANDARD: Provided as 20.0 ug of recombinant Human TGF Alpha. Quick-spin and reconstitute in 100 uL of sterile water. Further dilutions can be made in 0.05% Tween-20, 0.1% BSA in PBS.

DEVELOPING REAGENTS: Supplied with catalog # ending in "CKC".

- ELISA Coating/ Blocking Reagent (EA150C) 50.0 mL (5X Solution)
- Streptavidin-HRP (S100180C) 1.0 mL store @ -20 Deg. C.
- TMB Substrate Solutions Part A and Part B (50.0 mL each) cat # ES200C

HANDLING/ STORAGE: Reconstitute reagents when ready to build ELISA assay. Antibodies (Capture and Tracer) can be stored for approximately one month at 4 Degrees C. Or store frozen at -20 Degrees C. for up to 6 months. Standard (rec. Human TGF Alpha) can be stored in liquid state (@ 4 Deg. C.) For up to one week, or store frozen, with addition of 0.1% BSA, at -20 Deg. C. for up to 2 months. AVOID repeat freeze-thaw.

MATERIALS RECOMMENDED:

ELISA Microplates: Nunc Maxisorp, Prod. # 4420404 Tween -20.

BSA

Streptavidin-HRP: ANTIGENIX Cat no. S100180 or similar

TMB Substrate (ANTIGENIX cat # ES200)

Dubelco's PBS (10X)

ANTIGENIX ELISA Coating Stabilizer (cat no: EA150)

RECOMMENDED SOLUTIONS:

See ANTIGENIX Developing reagents above.

PBS: Dilute to 1XPBS in sterile water WASH BUFFER: 0.05% Tween-20 in PBS.

BLOCK BUFFER: use ANTIGENIX AMERICA coating stabilizer (EA150)

or 1% BSA in PBS

Substrate Solution: TMB Substrate Solution (ANTIGENIX # ES200)

Diluent: 0.05% Tween-20, 0.1% BSA in PBS

2N Sulfuric acid (stop solution)

PLATE PREPARATION:

- 1. Dilute **100 uL of** capture antibody with 0.05M Carbonate buffer (or PBS) to concentration 1.0 ug/mL. To optimize coating concentration on first test plate, use 2 test concentrations of capture antibody (for example: 1.0 ug/mL, and 0.5 ug/mL).
 - Immediately add 100 uL to each ELISA well. (With first test plate, coat one-half of plate with each concentration, and note which wells have different concentrations). Seal the plate and incubate overnight at room temperature.
- 2. Aspirate wells to remove all liquid and wash 4 times using 300 uL of wash buffer per well. After last wash, add 200 uL ANTIGENIX AMERICA ELISA coating stabilizer (cat # EA150) and incubate for 60 minutes at room temperature. (With coating stabilizer, DO NOT let plate dry prior to use of coating stabilizer. This will stabilize and Block in one step! Refer to data sheet EA150 for complete description of use.
- 3. With ANTIGENIX coating stabilizer (**recommended**) aspirate plate but **DO NOT WASH**. Dry plate in humidity controlled chamber or similar. (see data sheet EA150). With standard block reagent, aspirate plate and wash 3X with 300 uL wash buffer.

PROTOCOL:

STANDARD/SAMPLE: Dilute **a portion of the** standard (store unused standard in aliquots, high concentration, frozen -20 Deg. C.) from 20.0 ng/mL (**adjust**, depending on desired range, and sensitivity of first standard curve) to zero in diluent (serial dilution). Immediately add 100 uL of standard or sample to each well in duplicate. Incubate at room temp. for approx. 2 hours.

DETECTION: Aspirate and wash plate 4 times. Dilute portion of detection (Tracer) antibody in diluent to concentration of 0.20 μ mL. Add 100 μ mL. Incubate at room temperature for 2 hours. Note: detection antibody can be used in approximate range of 0.10 - 0.50 μ mL, you may need to optimize for subsequent plates.

STREPTAVIDIN-HRP: Aspirate and wash plate 4 times. Dilute Streptavidin-HRP conjugate approx. 1:2,000 in diluent (approx. 1-2 ug/mL concentration for cat # S100180). (follow recommended dilution of manufacturer). (May need to optimize) Add 100 uL per well, incubate 30 minutes at room temperature.

SUBSTRATE: Aspirate and wash plate 4 times. Add 100 uL substrate solution to each well. (follow directions from manufacturer) Incubate at room temp. for color development. Monitor color development with plate reader at 650 nm wavelength. (for blue color). Stop the color reaction after 10 - 30 minutes by adding 100 uL of 2N Sulfuric acid to each well. Then, read plate @ 450 nm after addition of stop solution.

NOTE: reliable standard curves are obtained when O.D. readings do not exceed 0.2 units for the zero standard concentration, or 2.2 units for the highest standard concentration. Monitor the plate every 5 minutes for approximately 30 minutes.

WARRANTY:

Products sold hereunder are warranted only to conform to the quantity and contents stated on the label at the time of delivery to the customer. There are no warranties, expressed or implied, which extend beyond the description on the label of the product.

RESEARCH USE ONLY -NOT For DIAGNOSTIC USE

ELISA Construction Kits <u>Developing/ Accessory Reagents Available</u>:

Streptavidin-HRP; S100180, 1.0 mL \$ 190.00 USD; suitable for 5,000 ELISA wells

TMB Substrate; ES200, 100 mL x 2, \$95.00; suitable for 2,000 ELISA wells

ELISA Coating Stabilizer; **EA150**; \$190.00, 100 mL (5X); suitable for 2,500 wells

Get All three reagents above as "ELISA Construction Pack"; EA700; \$410.00 USD.

ANTIGENIX AMERICA Inc.

WWW.ANTIGENIX.com