

ANTIGENIX AMERICA inc.

PRODUCT INFORMATION SUMMARY

Human IL-1RA ELISA Construction Kit

Product Number RHF718CK

Approx. 960 tests

Capture Antibody 100.0 ug

Biotin-Labeled tracer 25.0 ug

Antigen Standard 20.0 ug

Product Number **RHF718CKC**

With Developing Reagents:

ELISA Coating Stabilizer 50 mL

Streptavidin-HRP 0.5mL

TMB Substrate (50 mL x 2)

WASH Buffer (20X) 100 mL

RHF718CKP: with Developing reagents and 10 blank plates

DESCRIPTION:

Human IL-1RA ELISA CONSTRUCTION Kit provides antigen affinity purified capture and tracer antibodies, and antigen standard for development of **approximately** ten microplate assays.

Note: Reconstitute components only when ready to run assay.

CAPTURE ANTIBODY:

Provided as lyophilized, **100.0 ug**, additive-free. Reconstitute in 500 uL sterile water (200.0 ug/mL). (FREEZE aliquots for long-term storage)

TRACER ANTIBODY:

Provided as 25 ug of Biotin labeled, antigen-affinity purified antibody, additive-free. Reconstitute in 500 uL sterile water **containing 0.1% BSA**. (FREEZE aliquots for long-term storage)

STANDARD: Provided as 20.0 ug (**see vial**) of recombinant Human IL-1RA. Quick-spin and reconstitute in 50 uL of distilled water Further dilutions can be made in 0.1% BSA in PBS.

DEVELOPING REAGENTS: Supplied with catalog # ending in "CKC".

- ELISA Coating/ Blocking Reagent (EA150C) 50.0 mL (5X Solution)
- Streptavidin-HRP (S100180C) 0.5 mL - store @ -20 Deg. C.
- TMB Substrate Solutions - Part A and Part B (50.0 mL each) cat # ES200C
- WASH Buffer (20X)- Dilute 1 part with 19 parts distilled water

HANDLING/ STORAGE: Reconstitute reagents when ready to build ELISA assay. Antibodies (Capture and Tracer) can be stored for approximately one month at 4 Degrees C. Or store **frozen** at -20 Degrees C. for up to 6 months. Standard (rec. Human IP-10) can be stored in liquid state (@ 4 Deg. C.) For up to one week, or store **frozen, with addition of 0.1% BSA**, at -20 Deg. C. for up to 2 months. AVOID repeat freeze-thaw.

MATERIALS RECOMMENDED:

ELISA Microplates: Nunc Maxisorp, Prod. # 4420404
Tween -20.
BSA
Streptavidin-HRP: ANTIGENIX Cat no. **S100180** or similar
TMB Substrate : ANTIGENIX cat # **ES200** or similar
Dubelco's PBS (10X)
ANTIGENIX **ELISA Coating Stabilizer** (cat no: **EA150**)

RECOMMENDED SOLUTIONS:

See **ANTIGENIX Developing reagents above.**

PBS: Dilute to 1XPBS in sterile water
WASH BUFFER: ANTIGENIX **WB200** or 0.05% Tween-20 in PBS.
BLOCK BUFFER: **use ANTIGENIX AMERICA coating stabilizer (EA150)**
or 1% BSA in PBS
Substrate Solution: TMB Substrate Solution (cat # ES200)
Diluent: 0.1% BSA in PBS
2N Sulfuric acid (stop solution).

PLATE PREPARATION:

1. Dilute **portion** of capture antibody with 0.05M Carbonate buffer (or PBS) to concentration **1.0 ug/mL.**

Immediately add 100 uL to each ELISA well. Seal the plate and incubate overnight at room temperature.

2. Aspirate wells to remove all liquid and wash **4 times** using 300 uL of wash buffer per well. After last wash, add 200 uL ANTIGENIX AMERICA **ELISA coating stabilizer - recommended! -(cat # EA150)** and incubate for 60 minutes at room temperature. (With coating stabilizer, DO NOT let plate dry prior to use of coating stabilizer. This will **stabilize and Block in one step!** Refer to data sheet EA150 for complete description of use.
3. With ANTIGENIX coating stabilizer (**recommended**) aspirate plate but **DO NOT WASH.** For extended storage of plate- dry plate in humidity controlled chamber or similar. (see data sheet EA150). With standard block reagent, aspirate plate and wash 3X with 300 uL wash buffer.

PROTOCOL:

STANDARD/SAMPLE: Dilute **a portion of the** standard (store unused standard in aliquots, high concentration, frozen -20 Deg. C.) from **2.0 ng/mL** to zero in diluent (serial dilution). Immediately add 100 uL of standard or sample to each well in duplicate. **Incubate** at room temp. for approx. **90 minutes.**

DETECTION: Aspirate and wash plate 4 times. **Dilute** portion of detection (Tracer) antibody in diluent to concentration of 0.20 ug/mL. (1:250 from 50.0 ug/mL) **Add** 100 uL per well. **Incubate** at room temperature for 30 minutes - 1 hour. Note: detection antibody can be used in approximate range of 0.10 - 0.50 ug/mL, you may need to optimize for subsequent plates.

STREPTAVIDIN-HRP: Aspirate and wash plate 4 times. **Dilute** Streptavidin-HRP conjugate approx. **1:2,000** in diluent (follow recommended dilution of manufacturer). (May need to optimize) **Add** 100 uL per well, incubate 30 minutes at room temperature.

SUBSTRATE: Aspirate and wash plate 4 times. **Add** 100 uL substrate solution to each well (follow directions for substrate preparation and handling from manufacturer's data sheet). Incubate at room temp. for color development. **Stop** the color reaction after 10 - 15 minutes by adding 100 uL of 2N Sulfuric acid to each well. Then, **read plate @ 450 nm** after addition of stop solution.

NOTE: reliable standard curves are obtained when O.D. readings do not exceed 0.25 units for the zero standard concentration, or 2.0 units for the highest standard concentration.

NOTE: Kit can be purchased with developing reagents, cat # **RHF718CKC** -
or with Developing reagents and ten blank ELISA plates:
RHF718CKP (see page 1.)

RESEARCH USE ONLY -NOT For DIAGNOSTIC USE

ANTIGENIX AMERICA Inc.

WWW.ANTIGENIX.COM