

## PRODUCT INFORMATION SUMMARY

### Human sTRAIL ELISA Construction Kit

Product Number **RHF760CKC**  
With Developing Reagents:

<b>Capture Antibody</b>	<b>100.0 ug</b>	ELISA Coating Stabilizer	50 mL
Biotin-Labeled tracer	25.0 ug	Streptavidin-HRP	0.5mL
Antigen Standard	1.0 ug or 10.0ug	TMB Substrate ( 50 mL x 2)	
		WASH Buffer (20X)	100 mL

**RHF760CKP:** w/ Developing Reagents & 10 ELISA plates

#### DESCRIPTION:

This ELISA CONSTRUCTION Kit provides antigen affinity purified capture and tracer antibodies, and antigen standard sufficient for **approximately** ten microplates.

Working concentrations must be optimized by customer.

Note: Reconstitute components only when ready to run assay.

#### CAPTURE ANTIBODY:

Provided as lyophilized, 100 ug, additive-free. Reconstitute in 0.50 mL sterile or distilled water (200.0 ug/mL).

#### TRACER ANTIBODY:

Provided as 25 ug ( lyophilized or as 0.5 mL liquid- **see vial**) of Biotin labeled, antigen-affinity purified antibody, additive-free. Reconstitute lyophilized vial in 500 uL sterile water **containing 0.1% BSA.**

For biotin tracer provided as 0.5 mL liquid- STORE refrigerated.

**STANDARD:** Provided as 10.0 ug of recombinant Human TRAIL. **Quick-spin** and reconstitute in 100 uL of sterile water. Further dilutions can be made in 0.1% BSA in PBS.

**HANDLING/ STORAGE:** Reconstitute reagents when ready to build ELISA assay. Antibodies (Capture and Tracer) can be stored for approximately one month at 4 Degrees C. Or store **frozen** at -20 Degrees C. for up to 6 months. Standard ( rec. Human TRAIL) can be stored in liquid state ( @ 4 Deg. C.) For up to one week, or store **frozen** at -20 Deg. C. for up to 2 months. AVOID repeat freeze-thaw.

**MATERIALS RECOMMENDED:**

ELISA Microplates: Nunc Maxisorp, Prod. # 4420404  
Tween -20.  
BSA  
Streptavidin-HRP: ANTIGENIX Cat no. **S100180** or similar  
TMB Substrate (ANTIGENIX cat # ES200)  
Dubelco's PBS (10X)  
ANTIGENIX **ELISA Coating Stabilizer** ( cat no: **EA150**)

**RECOMMENDED SOLUTIONS:**

PBS: Dilute to 1XPBS in sterile water  
WASH BUFFER: 0.05% Tween-20 in PBS.  
BLOCK BUFFER: **use ANTIGENIX AMERICA coating stabilizer (EA150)**  
**or** 1% BSA in PBS  
Substrate Solution: TMB Substrate Solution (ANTIGENIX # ES200)  
Diluent: 0.05% Tween-20, 0.1% BSA in PBS

**PLATE PREPARATION:**

1. **Dilute portion** of capture antibody with 0.05M Carbonate buffer (or PBS) to concentration **1.0 ug/mL**. Immediately **add** 100 uL to each ELISA well. Seal the plate and incubate overnight at room temperature.
2. Aspirate wells to remove all liquid and wash **4 times** using 300 uL of wash buffer per well. After last wash, add 200 uL ANTIGENIX AMERICA **ELISA coating stabilizer (cat # EA150)-recommended!** and incubate for 60 minutes at room temperature. ( With coating stabilizer, **DO NOT** let plate dry prior to use of coating stabilizer. This will **stabilize and Block in one step!** Refer to data sheet EA150 for complete description of use.
3. With ANTIGENIX coating stabilizer ( **recommended** ) aspirate plate but **DO NOT WASH**. For extended storage -Dry plate in humidity controlled chamber or similar. ( see data sheet EA150). With standard block reagent, aspirate plate and wash 3X with 300 uL wash buffer.

**PROTOCOL:**

**STANDARD/SAMPLE:** Dilute **a portion of the** standard ( store unused standard in aliquots, high concentration, frozen -20 Deg. C.) from 2.0 ng/mL (1:2, 8 point serial dilution). Immediately **add** 100 uL of standard or sample to each well in duplicate. **Incubate** at room temp. for approx. 90 minutes.

**DETECTION:** Aspirate and wash plate 4 times. **Dilute** detection (Tracer) antibody in diluent to concentration of 0.20 ug/mL. ( approx. 1: 250 from 50.0 ug/mL) **Add** 100 uL per well. Incubate at room temperature for 1 hour.

Note: detection antibody can be used in approximate range of 0.15 - 0.30 ug/mL, you may need to optimize for subsequent plates.

**STREPTAVIDIN-HRP:** Aspirate and wash plate 4 times. Dilute Streptavidin-HRP conjugate approx. 1:2000 in diluent ( follow recommended dilution of manufacturer). (May need to optimize) **Add** 100 uL per well, **incubate** 30 minutes at room temperature.

**SUBSTRATE:** Aspirate and wash plate 4 times. Add 100 uL substrate solution to each well. ( follow directions from manufacturer) **Incubate** at room temp. for color development. STOP color developmnet after approx. 10-15 minutes. Read plate at 450 nm within 30 minutes.

NOTE: reliable standard curves are obtained when O.D. readings do not exceed 0.30 units for the zero standard concentration, or 1.5 units for the highest standard concentration.

**RESEARCH USE ONLY -NOT For DIAGNOSTIC USE**

ANTIGENIX AMERICA Inc.

[WWW.ANTIGENIX.COM](http://WWW.ANTIGENIX.COM)