## ANTIGENIX AMERICA Inc.

# PRODUCT INFORMATION SUMMARY

"SUPER X" ELISA Kit 5 Plate Kit ( 5 x 96 tests)

Rat SDF-1 Alpha

Product Number RRF440CKX

Product Number RRF440CKX2 (2 x 96 tests)

RRF440CKX1 ( 96 tests)

# Items Provided:

PRE-COATED ELISA Plates (5 plates, 2 plates, or single plate)

Biotin-Labeled tracer 25.0 ug Antigen Standard 2.0 ug Streptavidin-HRP 0.5 mL

TMB Substrate  $25 \text{ mL} \times 2 \text{ (15.0 mL} - 2 \text{ plate)}$ 

Wash Buffer Concentrate 125 mL (20X); or 50 mL (2 plate kit)

#### **DESCRIPTION:**

Rat SDF-1 Alpha (CKCL12) **SUPER X ELISA Kit** provides **PRE-COATED** microplates (coated with antigen affinity purified capture antibody stabilized with our proprietary ELISA coating/blocking reagent. A biotin labeled tracer antibody, antigen standard, HRP developing reagents and wash buffer are included.

Note: Reconstitute components only when ready to run assay.

#### TRACER ANTIBODY:

Provided as 25 ug (lyophilized) or as 0.5 mL liquid

@ 50.0 ug/mL ( see vial) of Biotin labeled, antigen-affinity purified antibody, additive-free.

Reconstitute the 25 ug (lyophilized vial) in 500 uL sterile water containing 0.1% BSA. (FREEZE aliquots for long-term storage).

\*\*For liquid vial store refrigerated only ( contains preservative) \*\*.

STANDARD: Provided as 2.0 ug of recombinant Rat SDF-1 Alpha. Quick-spin and reconstitute in distilled water (pH 8.0) - concentration approx. 0.1 mg/mL. Further dilutions can be made in diluent provided or 0.1% BSA in PBS.

# **DEVELOPING REAGENTS:**

- Streptavidin-HRP ( S100180CX) 0.5 mL store @ -20 Deg. C.
- TMB Substrate Solutions Part A and Part B ( 25.0 mL each) cat # ES200CX
- Wash Buffer Concentrate (20X concentrate) Mix 1 volume of wash buffer with 19 volumes of distilled water. Stable for one month @ 4 Deg. C. once mixed to working volume. Concentrate may be stored at room temp.
- Diluent (10X) -Dilute 1 part with 9 parts distilled water
- STOP Solution (1X) Caution Caustic

HANDLING/ STORAGE: Reconstitute reagents when ready to build ELISA assay. Tracer Antibody be stored for approximately one month at 4 Degrees C. Or store frozen at -20 Degrees C. for up to 6 months. Standard (rec. SDF-1 alpha) can be stored in liquid state (@ 4 Deg. C.) For up to one week, or store @ -20 Deg. C. - with addition of 0.1% BSA, at -20 Deg. C. for up to 2 months. AVOID repeat freeze-thaw. Pre-coated ELISA plates should be stored in sealed plastic bag with desiccant pack.

Store Streptavidin-HRP (S100180XH) @ -20 Deg. C.

Store TMB solutions and wash buffer concentrate at 4 Deg. C. CAUTION: Substrate Solution B contains 20% acetone. FLAMMABLE. Keep away from sources of heat or flame.

# MATERIALS RECOMMENDED:

Tween -20.
BSA (ELISA grade only, protease-free)
Dubelco's PBS (10X)
PBS: Dilute to 1XPBS in sterile water

#### PLATE PREPARATION:

Plates are **pre-coated** with capture antibody and blocked/stabilized with ANTIGENIX proprietary ELISA coating stabilizer (EA150) and are **ready to use**!

Store plates refrigerated ( @ 4 Deg. C.) in sealed plastic bags with desiccant pack.

## PROTOCOL:

STANDARD/SAMPLE: Dilute a portion of the standard (store unused standard in aliquots, high concentration, frozen -20 Deg. C. with addition of 0.1% BSA) from 5.0 ng/mL to zero in diluent (1:2, 8-point serial dilution). Immediately add 100 uL of standard or sample to each well in duplicate. Incubate at room temp. for 90 minutes.

**DETECTION:** Aspirate and wash plate 4 times. **Dilute** portion of detection (Biotin Tracer) antibody in diluent to **concentration** of **0.20 ug/mL**. (Approx. 1:250 from 50.0 ug/mL) Add 100 uL per well. **Incubate** at room temperature for **50 minutes**.

STREPTAVIDIN-HRP: Aspirate and wash plate 4 times. Dilute Streptavidin-HRP conjugate approx. 1:2,000 in diluent. (May need to optimize) Add 100 uL per well, incubate 30 minutes at room temperature.

SUBSTRATE: Prepare substrate solution no more than 15 minutes before last incubation of assay: Mix one part TMB Solution A with one part TMB Solution B in a clean container. If, upon mixing, TMB solution turns blue - TMB solution is contaminated-DO NOT USE. Use mixed substrate solution WITHIN 2 Hours, and AVOID DIRECT LIGHT.

Aspirate and wash plate 4 times. Note: Wash steps are critical! Add 100 uL substrate solution to each well. Incubate at room temp. for color development. Add 100 uL of Stop solution ( 2N Sulfuric Acid) within 10 minutes to stop color development gently tap plate to mix. Read plate at 450 nm within 30 minutes of addition of stop solution.

NOTE: reliable standard curves are obtained when O.D. readings do not exceed 0.30 units for the zero standard concentration, or 2.0 units for the highest standard concentration.

RESEARCH USE ONLY -NOT For DIAGNOSTIC USE

ANTIGENIX AMERICA Inc. WWW.ANTIGENIX.com